

CLIENT INFORMATION

Client: *****

Requested On: Jun 25, 2025

Phone: *****

Email: *****

Kitting, Logistics, and Support provided by: SimpleLab, Inc.

Questions? For fastest assistance:

hello@gosimplelab.com

Do not contact facility technicians directly.

TESTING PERFORMED

Testing Requested: PFAS Water Test

Matrix: Drinking Water

Testing / Report ID: 24Z85H

Testing Facility: EMSL Analytical Inc

Facility Location: 200 Route 130 North
Cinnaminson, New Jersey 08077

SAMPLE INFORMATION

Collection Date: Jul 7, 2025

Collected By: *****

Received Date: Jul 8, 2025

Reported On: Jul 11, 2025

Sample Location: Fiji (in plastic)

Sample Address: *****

TESTING NOTES FROM SIMPLELAB

There were no problems with analytical events associated with this report unless noted. Quality control data is within laboratory defined or method specified acceptance limits except where noted. If you have any questions regarding these test results, please contact hello@gosimplelab.com

TEST RESULTS

ANALYTE	UNIT	RESULT	MDL	RL	METHOD
N-ethyl perfluorooctanesulfonamidoacetic acid	µg/L	NOT DETECTED	0.00075	0.00193	EPA 537.1
N-methyl perfluorooctanesulfonamidoacetic acid	µg/L	NOT DETECTED	0.00071	0.00193	EPA 537.1
Perfluorobutanesulfonic acid	µg/L	NOT DETECTED	0.00062	0.00193	EPA 537.1
Perfluorodecanoic acid	µg/L	NOT DETECTED	0.00074	0.00193	EPA 537.1
Perfluorododecanoic acid	µg/L	NOT DETECTED	0.001	0.00193	EPA 537.1
Perfluoroheptanoic acid	µg/L	NOT DETECTED	0.00078	0.00193	EPA 537.1
Perfluorohexanesulfonic acid	µg/L	NOT DETECTED	0.00062	0.00193	EPA 537.1
Perfluorohexanoic acid	µg/L	0.000878	0.0007	0.00193	EPA 537.1
Perfluorononanoic acid	µg/L	NOT DETECTED	0.00071	0.00193	EPA 537.1
Perfluorooctanesulfonic acid	µg/L	NOT DETECTED	0.00057	0.00193	EPA 537.1
Perfluorooctanoic acid	µg/L	NOT DETECTED	0.00072	0.00193	EPA 537.1
Perfluorotetradecanoic acid	µg/L	NOT DETECTED	0.00133	0.00193	EPA 537.1
Perfluorotridecanoic acid	µg/L	NOT DETECTED	0.00116	0.00193	EPA 537.1
Perfluoroundecanoic acid	µg/L	NOT DETECTED	0.00071	0.00193	EPA 537.1

How To Read Your SimpleLab PDF Report

MDL: The Method Detection Limit is the lowest concentration which the analysis team and instrumentation is configured to measure.

Use of this report can only be done in full, with no alterations or additions. SimpleLab maintains the right to enforce this and by accepting this report you agree to abide by this policy.

RL: The Method Reporting Limit is the lowest level at which the laboratory can confidently and accurately quantify the concentration.