

CLIENT INFORMATION

Client: *****
 Requested On: Jun 25, 2025
 Phone: *****
 Email: *****

Kitting, Logistics, and Support provided by: SimpleLab, Inc.

Questions? For fastest assistance:
hello@gosimplelab.com
 Do not contact facility technicians directly.

TESTING PERFORMED

Testing Requested: PFAS Water Test
 Matrix: Drinking Water
 Testing / Report ID: PFKXQE
 Testing Facility: EMSL Analytical Inc
 Facility Location: 200 Route 130 North
 Cinnaminson, New Jersey 08077

SAMPLE INFORMATION

Collection Date: Jul 20, 2025
 Collected By: *****
 Received Date: Jul 23, 2025
 Reported On: Jul 25, 2025
 Sample Location: Voss (in Glass)
 Sample Address: *****

TESTING NOTES FROM SIMPLELAB

There were no problems with analytical events associated with this report unless noted. Quality control data is within laboratory defined or method specified acceptance limits except where noted. If you have any questions regarding these test results, please contact hello@gosimplelab.com

TEST RESULTS

ANALYTE	UNIT	RESULT	MDL	RL	METHOD
N-ethyl perfluorooctanesulfonamidoacetic acid	µg/L	NOT DETECTED	0.00069	0.00177	EPA 537.1
N-methyl perfluorooctanesulfonamidoacetic acid	µg/L	NOT DETECTED	0.00065	0.00177	EPA 537.1
Perfluorobutanesulfonic acid	µg/L	NOT DETECTED	0.00057	0.00177	EPA 537.1
Perfluorodecanoic acid	µg/L	NOT DETECTED	0.00068	0.00177	EPA 537.1
Perfluorododecanoic acid	µg/L	NOT DETECTED	0.00092	0.00177	EPA 537.1
Perfluoroheptanoic acid	µg/L	NOT DETECTED	0.00072	0.00177	EPA 537.1
Perfluorohexanesulfonic acid	µg/L	NOT DETECTED	0.00057	0.00177	EPA 537.1
Perfluorohexanoic acid	µg/L	NOT DETECTED	0.00064	0.00177	EPA 537.1
Perfluorononanoic acid	µg/L	NOT DETECTED	0.00065	0.00177	EPA 537.1
Perfluorooctanesulfonic acid	µg/L	NOT DETECTED	0.00053	0.00177	EPA 537.1
Perfluorooctanoic acid	µg/L	NOT DETECTED	0.00066	0.00177	EPA 537.1
Perfluorotetradecanoic acid	µg/L	NOT DETECTED	0.00122	0.00177	EPA 537.1
Perfluorotridecanoic acid	µg/L	NOT DETECTED	0.00106	0.00177	EPA 537.1
Perfluoroundecanoic acid	µg/L	NOT DETECTED	0.00065	0.00177	EPA 537.1

How To Read Your SimpleLab PDF Report

MDL: The Method Detection Limit is the lowest concentration which the analysis team and instrumentation is configured to measure.

Use of this report can only be done in full, with no alterations or additions. SimpleLab maintains the right to enforce this and by accepting this report you agree to abide by this policy.

RL: The Method Reporting Limit is the lowest level at which the laboratory can confidently and accurately quantify the concentration.

Use of this report can only be done in full, with no alterations or additions. SimpleLab maintains the right to enforce this and by accepting this report you agree to abide by this policy.